

## Declaration of conformity

We,

**GDF Gesellschaft für dentale Forschung und Innovationen GmbH**  
(name of producer)

Dieselstr. 5-6  
61191 Rosbach/Germany  
(address)

declare in our own responsibility that the product/product-line:

### **Fixing Composite / Root Pin Cement**

risk class **Ila** according to annex IX of the guideline 93/42 EC  
(classification rule: 8)

which this declaration is related to, is in accordance with the following  
conformity assessment procedure:

**annex II without chapter 4**  
(list the according annexes II - VII)

of the guideline 93/42/EC of June 14,  
1993, on Medical Devices.

This declaration is valid for all products supplied until October 11, 2023.

Rosbach, October 06, 2020

(place and date of issue)



Dr. Uwe Krichbaum  
Safety Officer for Medical Devices

Notified Body: DQS Medizinprodukte GmbH, August-Schanz-Str. 21, 60433 Frankfurt -  
Germany  
CE-code: 0297